Patient:	ABED MILOUDA (female, #RF84225)
Series:	Dose SR (#10002)
Manufacturer:	UIH (uCT 530, #300183)
Preliminary Flag:	FINAL
Completion Flag:	COMPLETE
Verification Flag:	UNVERIFIED
Content Date/Time:	2026-11-11 15:28:21

X-Ray Radiation Dose Report

Concept Modifier: Procedure reported [Annex 1]
Observation Context: Observer Type = Device (121007, DCM)
Observation Context: Device Observer UID = 300183
Observation Context: Start of X-Ray Irradiation = 2026-11-11 15:26:12 +00:00
Observation Context: End of X-Ray Irradiation = 2026-11-11 15:26:44 +00:00
Observation Context: Scope of Accumulation [Annex 2]

Source of Dose Information:
Automated Data Collection (113856, DCM)

CT Accumulated Dose Data

Total Number of Irradiation Events:
2 {events}
CT Dose Length Product Total:
907.31 mGy.cm

CT Acquisition

Acquisition Protocol:
Scout
Target Region:
Head (T-D1100, SRT)
CT Acquisition Type:
Constant Angle Acquisition (113805, DCM)
Irradiation Event UID:
1.2.156.112605.137174094783624.261111152554.9.7040.39842

CT Acquisition Parameters

Exposure Time:
1.861918 s
Scanning Length:
252.1473 mm
Nominal Single Collimation Width:
0.55 mm
Nominal Total Collimation Width:
2.2 mm

Number of X-Ray Sources:
1 {X-Ray sources}

CT X-Ray Source Parameters

Identification of the XRay Source:
A
KVP:
120 kV
Maximum X-Ray Tube Current:
87.77072 mA
X-Ray Tube Current:
80 mA

CT Dose

Mean CTDIvol:
0.180974 mGy
CTDIw Phantom Type:
IEC Head Dosimetry Phantom (113690, DCM)
DLP:
4.57 mGy.cm

CT Acquisition

Acquisition Protocol:
Head Helical
Target Region:
Head (T-D1100, SRT)
CT Acquisition Type:
Spiral Acquisition (P5-08001, SRT)
Irradiation Event UID:
1.2.156.112605.137174094783624.261111152554.9.7040.69842

CT Acquisition Parameters

Exposure Time:
10.36195 s
Scanning Length:
169.4148 mm
Nominal Single Collimation Width:
0.55 mm
Nominal Total Collimation Width:
22 mm
Pitch Factor:
0.675 {ratio}
Number of X-Ray Sources:
1 {X-Ray sources}

CT X-Ray Source Parameters

Identification of the XRay Source:

A
KVP:
120 kV
Maximum X-Ray Tube Current:
253.7046 mA
X-Ray Tube Current:
251.4375 mA
Exposure Time per Rotation:
0.8 s

CT Dose

Mean CTDIvol:
46.8757 mGy
CTDIw Phantom Type:
IEC Head Dosimetry Phantom (113690, DCM)
DLP:
902.74 mGy.cm

Annex

Annex 1

Procedure reported:
Computed Tomography XRay (P5-08000, SRT)

Concept Modifier: Has Intent = Diagnostic Intent (R-408C3, SRT)

Annex 2

Scope of Accumulation:
Study (113014, DCM)

Properties: Study Instance UID = 1.2.840.4892943.343.20261111031954.3832.92548

This page was generated from a DICOM Structured Reporting document by OFFIS DCMTK 3.6.5.