Patient:	NAIMA OUABTAR (female, *1981-11-13, #RF84292)
Series:	Dose SR (#10002)
Manufacturer:	UIH (uCT 530, #300183)
Preliminary Flag:	FINAL
Completion Flag:	COMPLETE
Verification Flag:	UNVERIFIED
Content Date/Time:	2026-11-13 09:33:08

X-Ray Radiation Dose Report

Concept Modifier: Procedure reported [Annex 1]
Observation Context: Observer Type = Device (121007, DCM)
Observation Context: Device Observer UID = 300183
Observation Context: Start of X-Ray Irradiation = 2026-11-13 09:29:55 +00:00
Observation Context: End of X-Ray Irradiation = 2026-11-13 09:33:05 +00:00
Observation Context: Scope of Accumulation [Annex 2]

Source of Dose Information:
Automated Data Collection (113856, DCM)

CT Accumulated Dose Data

Total Number of Irradiation Events:
3 {events}
CT Dose Length Product Total:
987.38 mGy.cm

CT Acquisition

Acquisition Protocol:
Scout
Target Region:
Abdomen (T-D4000, SRT)
CT Acquisition Type:
Constant Angle Acquisition (113805, DCM)
Irradiation Event UID:
1.2.156.112605.137174094783624.261113092345.9.7040.37087

CT Acquisition Parameters

Exposure Time:
4.108235 s
Scanning Length:
702.0408 mm
Nominal Single Collimation Width:
0.55 mm
Nominal Total Collimation Width:
2.2 mm

Number of X-Ray Sources:
1 {X-Ray sources}

CT X-Ray Source Parameters

Identification of the XRay Source:
A
KVP:
120 kV
Maximum X-Ray Tube Current:
91.51604 mA
X-Ray Tube Current:
40 mA

CT Dose

Mean CTDIvol:
0.03856682 mGy
CTDIw Phantom Type:
IEC Body Dosimetry Phantom (113691, DCM)
DLP:
2.71 mGy.cm

CT Acquisition

Acquisition Protocol:
Arterial Phase
Target Region:
Abdomen (T-D4000, SRT)
CT Acquisition Type:
Spiral Acquisition (P5-08001, SRT)
Irradiation Event UID:
1.2.156.112605.137174094783624.261113092345.9.7040.95614

CT Acquisition Parameters

Exposure Time:
19.27214 s
Scanning Length:
627.4565 mm
Nominal Single Collimation Width:
0.55 mm
Nominal Total Collimation Width:
22 mm
Pitch Factor:
1.075 {ratio}
Number of X-Ray Sources:
1 {X-Ray sources}

CT X-Ray Source Parameters

Identification of the XRay Source:

A
KVP:
120 kV
Maximum X-Ray Tube Current:
260.0961 mA
X-Ray Tube Current:
165.8571 mA
Exposure Time per Rotation:
0.7 s

CT Dose

Mean CTDIvol:
9.074769 mGy
CTDIw Phantom Type:
IEC Body Dosimetry Phantom (113691, DCM)
DLP:
590.91 mGy.cm

CT Acquisition

Acquisition Protocol:
Portal Phase
Target Region:
Abdomen (T-D4000, SRT)
CT Acquisition Type:
Spiral Acquisition (P5-08001, SRT)
Irradiation Event UID:
1.2.156.112605.137174094783624.261113092346.9.7040.38844

CT Acquisition Parameters

Exposure Time:
12.69922 s
Scanning Length:
405.4554 mm
Nominal Single Collimation Width:
0.55 mm
Nominal Total Collimation Width:
22 mm
Pitch Factor:
1.075 {ratio}
Number of X-Ray Sources:
1 {X-Ray sources}

CT X-Ray Source Parameters

Identification of the XRay Source:
A
KVP:
120 kV
Maximum X-Ray Tube Current:

244.8095 mA
X-Ray Tube Current:
167.3929 mA
Exposure Time per Rotation:
0.7 s

CT Dose

Mean CTDIvol:
9.164591 mGy
CTDIw Phantom Type:
IEC Body Dosimetry Phantom (113691, DCM)
DLP:
393.76 mGy.cm

Annex

Annex 1

Procedure reported:
Computed Tomography XRay (P5-08000, SRT)

Concept Modifier: Has Intent = Diagnostic Intent (R-408C3, SRT)

Annex 2

Scope of Accumulation:
Study (113014, DCM)

Properties: Study Instance UID = 1.2.840.4892943.343.20261113091846.8146.92619

This page was generated from a DICOM Structured Reporting document by OFFIS DCMTK 3.6.5.